

# EC - DECLARATION OF CONFORMITY

In compliance with (EU) 2017/745 Medical Device Regulation

|                               |                                                                    |
|-------------------------------|--------------------------------------------------------------------|
| <b>Product name</b>           | Medlogics Anti-Decubitus Seat Cushion                              |
| <b>REF</b>                    | AC100                                                              |
| <b>Basic UDI-DI</b>           | 3800501345AC100BZ                                                  |
| <b>MP Class and Rule</b>      | Class I, According to Rule 1, Annex VIII, Regulation (EU) 2017/745 |
| <b>UMDNS</b>                  | 15214                                                              |
| <b>Standards</b>              | BS 5852 - Crib 5                                                   |
| <b>Conformity Assessments</b> | Annex II+III of Regulation (EU) 2017/745                           |

We herewith declare that the above mentioned products meet the requirements of Medical Device Regulation (EU) 2017/745 and the applicable standards above Regulations by bearing CE-marking.

Version: 1  
Date: September 2021  
Valid through: September 2026

Department: Regulatory Affairs  
Name: Stanimir Nenov  
Signature: *S Nenov*

## Headquarters:

Medlogics.com LTD  
Lyuben Karavelov 6  
9000 Varna, Bulgaria

**E** info@medlogics.com  
**W** www.medlogics.com

